

Announcement No. _____

V6AI- AM/KUTE-FM Radio Announcement Form
Department of Youth & Civics Affairs

To: Director, DYCA
From: Dpt. Of Health Services

Starting Date: 5/2/25

Ending Date: 6/20/25

Frequency:

☒ Regular – 3x per day ☐ Emergency – ASAP

Languages (*check all that apply*):

☒ English ☒ Yapese ☒ Outer Island

Directions:

All announcements should be type-written and double spaced. If more space is needed, please attach an extra sheet.

PLEASE REFER TO ATTACHMENT

Reviewed & Approved By: _____

DYCA, Director Julius Alex Gilfiley

5/7/25

Date

FOR V6AI-AM/KUTE-FM OFFICIAL USE ONLY

Received By: _____

Acting Chief of Media

Date: _____

Time: _____



The State of Yap
Department of Health Services
P.O. Box 148, Colonia, Yap 96943
Federated States of Micronesia
Telephone #: 011 (691) 350-2110



Request For Quotation:

The Department of Health Service is tendering its Request for Quotation (RFQ) and will be accepting sealed bids for **Yap State Hospital Medical Equipment**, which specifications are listed below, beginning on May 02, 2025 until at 3:00 pm at local standard time on June 20, 2025:

Medical Equipment List Including specifications:

1. Traction Bed with Mattress
2. Hospital Manual Bed with Mattress
3. Video Laryngoscope
4. Blood/Fluid Warmer
5. Fiberoptic Bronchoscope for Adult
6. Fiberoptic Bronchoscope for Pediatric
7. Multi-Channel Infusion Pump System

All sealed bids shall be addressed to the Director of Health Services, P.O. Box 148, Colonia, Yap State 96943 and shall be labeled **"Bid Request for Quotation for Yap State Hospital Medical Equipment"** and shall bear the name and address of the bidder on the outside of the envelope. Bid request thus received will be publicly opened and read by the Director or her designee at 3:00 pm local time on June 20, 2025.

The Department of Health Services reserves the right to reject any and all of the bids submitted. The award of contract, if awarded, will be made to the lowest responsible bidder whose proposal complies with all the requirements prescribed, in accordance with the Yap State Bidding Regulations.

Bid announcement can also be viewed at the Yap State Government Site.
<http://www.yapstategov.org/>

Please direct any questions or inquiries to Bobby Nepomuceno or Anthony Bay at 350-2110.



The State of Yap
Department of Health Services
P.O. Box 148, Colonia, Yap 96943
Federated States of Micronesia
Telephone #: 011 (691) 350-2110



SPECIFICATIONS FOR BLOOD/FLUID WARMER

1. Electrical: 110 Volts, 60 Hertz
2. LCD Display
3. Temperature Range: 32-42 Degrees Centigrade
4. With Pole Clamp
5. With Separate Service Manual (Soft and Hard Copy)
6. With Operation Manual
7. With shipping and handling cost
8. With delivery lead time of equipment
9. CIF-YAP



The State of Yap
Department of Health Services
P.O. Box 148, Colonia, Yap 96943
Federated States of Micronesia
Telephone #: 011 (691) 350-2110



SPECIFICATIONS FOR AN INFUSION PUMP SYSTEM

1. Electrical: 110 Volts, 60 Hertz
2. Battery-Operated, Minimum of 4 Hrs of Continuous Operation
3. Can Infuse I.V. Fluids and Drugs
4. Multi-Channel Pump Modules Delivery System
5. Includes 60pcs. Start-Up Infusion Sets
6. Includes Service Manual (Soft and Hard Copy)
7. Includes Operations Manual
8. Includes Shipping Cost
9. With Delivery Lead Time of Equipment
10. CIF-YAP



The State of Yap
Department of Health Services
P.O. Box 148, Colonia, Yap 96943
Federated States of Micronesia
Telephone #: 011 (691) 350-2110



Specifications for Orthopedic Traction Bed

1. Designed for Orthopedic surgeries and traumatic treatment backrest and leg rest adjustment capabilities.
2. Capable to perform expanding traction to different parts of vertebra and extremities backrest lifting angle 0°-70°15°
3. Foot-rest lifting angle 0°-30°25° Load Bearing 240 kg
4. Height adjustment 500mm (from floor)
5. Dimension 2020x900mm
6. With shipping and handling cost
7. With delivery lead time of equipment
8. CIF-Yap



The State of Yap
Department of Health Services
P.O. Box 148, Colonia, Yap 96943
Federated States of Micronesia
Telephone #: 011 (691) 350-2110



Specification of a Manual Hospital Bed

1. Frame Features:

- ✓ Standard Size
- ✓ 4 Manual Crank Functions (Backrest tilt angle is 0-80 degrees; Knee-break tilt angle is 0-40 degrees; Height adjustable 350-700mm; Trendelenburg is 12 to 30 degrees; Reversed Trendelenburg is 12 to 30 degrees)
- ✓ Concealed type crank-handle system(foldable)
- ✓ Frame: Steel with Antibacterial Coating
- ✓ Support mattress platform ABS Molding with 4 sections
- ✓ Patient helper bar
- ✓ Permanently mounted Durable Plastic or Steel Head and Foot Board: no sawdust filling materials
- ✓ Four-folding small PP side rails (not removable)
- ✓ IV Socket: Minimum 2 each side
- ✓ Static Load or Capacity: 240kgs(512lbs)
- ✓ Lockable Caster 125mm(diameter)

2. Mattress Features:

- ✓ Standard size mattress (Mattress appropriate as per bed size offer)
- ✓ 6 inches thick mattress
- ✓ Durable, water proof, antimicrobial material covering
- ✓ Anti -static, anti-fungal, moisture permeability and flame-retardant cover
- ✓ Antibacterial, anti-static, acid resistant and waterproof
- ✓ High density and flame retardant foam
- ✓ Pressure release in prevention of pressure ulcers

3. Others:

- ✓ Includes additional 6 pcs. mattress cost
- ✓ Includes service Manual
- ✓ Includes one year warranty
- ✓ Includes pictures of manual bed and mattress
- ✓ Includes Delivery Lead Time of Equipment
- ✓ Quotation must be CIF-Yap



The State of Yap
Department of Health Services
P.O. Box 148, Colonia, Yap 96948
Federated States of Micronesia
Telephone #: 011 (691) 350-2110



BID SPECIFICATION FOR A FIBREOPTIC BRONCHOSCOPE (Adult)

1. The bronchoscope shall be forward viewing and the optical system shall be capable of Providing;
 - a) a field of view of at least 100°
 - b) a depth of field adjustable from 3-50 mm.
2. The bending angle range of the tip shall be;
 - a) Up - 180° , Down - 130°
3. The outer diameter of the working section shall be not greater than 6 mm and the working length shall be not less than 540 mm
4. Each unit shall be supplied with a Halogen or incandescent or LED Light source with built in spare lamp for quick changeover.
5. Each unit shall be supplied with the following standard accessories;

a) Biopsy forceps (Fenestrated)	- 02 Nos.
b) Biopsy needles (One with a center spike)	- 02 Nos.
c) Cytology brushes	- 01 Nos.
d) Brush for channel cleaning	- 01 Nos.
e) Mouthpiece (bite guard)	- 02 Nos.
f) Lens cleaner	- 01 Nos.
g) Cleaning brushes	- 04 Nos.
h) Foreign Body removal snare	- 02 Nos.
i) Grasping Forceps	- 02 Nos.
6. A list of optional accessories (eg.: different types of biopsy forceps etc.) together with their unit cost shall be quoted separately.
7. The applicable sterilization technique shall be clearly stated.
8. With service manual, soft and hard copy
9. With operations manual, soft and hard copy
10. With installation & commissioning
11. With warranty:

- i. 2 years of parts and labor
- ii. 2 preventive maintenance visits

12. With shipping & handling costs

13. With delivery lead time of equipment

14. CIF- YAP



The State of Yap
Department of Health Services
P.O. Box 148, Colonia, Yap 96943
Federated States of Micronesia
Telephone #: 011 (691) 350-2110



BID SPECIFICATION FOR A FIBREOPTIC BRONCHOSCOPE (Pediatric)

1. The bronchoscope shall be forward viewing and the optical system shall be capable of Providing;
 - a) A field of view of at least 110°
 - b) A depth of field adjustable from 2-50 mm.
2. The bending angle range of the tip shall be;
 - a) Up - 180°, Down - 130°
3. The outer diameter of the working section shall be not greater than 4.1 mm and the Working length shall be not less than 360 mm
4. Each unit shall be supplied with a Halogen or incandescent or LED Light source with Built in spare lamp for quick changeover.
5. Each unit shall be supplied with the following standard accessories;

a) Biopsy forceps (Fenestrated)	- 02 Nos.
b) Biopsy needles (One with a center spike)	- 02 Nos.
c) Cytology brushes	- 01 Nos.
d) Brush for channel cleaning	- 01 Nos.
e) Mouthpiece (bite guard)	- 02 Nos.
f) Lens cleaner	- 01 Nos.
g) Cleaning brushes	- 04 Nos.
h) Foreign Body removal snare	- 02 Nos.
i) Grasping Forceps	- 02 Nos.
06. A list of optional accessories (eg.: different types of biopsy forceps etc.) together with Their unit cost shall be quoted separately.
07. The applicable sterilization technique shall be clearly stated.
08. With service manual, soft and hard copy
09. With operations manual, soft and hard copy
10. With installation & commissioning

11. With warranty:

- a. 2 years of parts and labor
- b. 2 preventive maintenance visits

12. With shipping & handling costs

13. With delivery lead time of equipment

14. CIF- YAP



The State of Yap
Department of Health Services
P.O. Box 148, Colonia, Yap 96943
Federated States of Micronesia
Telephone #: 011 (691) 350-2110



BID Specification for Video Laryngoscope

1. For viewing vocal folds and glottis. Surgical and mechanical ventilation / intubation
2. A light source on or via the blade illuminates the larynx to allow viewing and tube passage. The unit is handheld with internal batteries and has interchangeable, rigid blades of different sizes with display screen and external connection facility for demonstration.
3. Fiber optic Laryngoscope with LCD Display preferably should be single patient use to ensure no infection to the patients, should comprise of disposable handle and reusable light source using the latest LED technology.
4. The main body of the handle should incorporate an excellent grip & should feel even wearing a glove. There should be a freely moving light intensifier of light from the light source through to the tip of the fiber optic blade to prevent any possibility of cross contamination. The unit should allow the blade to be inserted easily & should provide a positive locking mechanism when moved in to the closed position. The patient contact material should be biocompatible.
5. Display of LCD screen size 3.5" Full View, Resolution Ratio 640x480 RGB, Aspect Ratio 4:3 Video Refresh Rate: 30FPS, illumination: LED. Screen is reinforced with ant shatter protection.
6. Resolution ratio 2.0 M Pixels, Field Angle 66°, luminance: *800LUX
7. Easy to Establish and store files Connect to External Monitor and convenient for training and presentation
8. Light weight, mobile, portable and
 - i. Handheld unit, single piece when in use
 - ii. on/off switch to be robust and easy to use
 - iii. External material to be non-ferrous
 - iv. Blades to be surgical grade **stainless steel** and autoclavable
 - v. Supplied in protective, enclosable case
9. Power Requirements is independent of external source
10. Battery Type: Rechargeable Lithium Battery Long working time :> 200 minutes Voltage 3.7 V Capacity: 3200mAh Battery Life Cycle: more than 300 charges Charging time: Less than 8 Hrs
11. Power Adapter: Type: Rechargeable Lithium Battery Long Working Time: > 200 minutes Voltage
 - i. 3.7 V Capacity: 3200 mAh Battery Life Cycle: More than 300 Charges charging time: Less than 8 Hrs
12. Accessories (mandatory, Batteries, light source, Macintosh blades of all available adult and Pediatric sizes standard,
13. Operating condition: Atmosphere / Ambiance (air conditioning, Capable of operating continuously in ambient temperature of 0 to 50 deg C and relative humidity of 15 to 90% in ideal circumstances. an ambient air velocity is less than 0.3 m/s. 6.1 humidity, dust...) Liquid splash resistant Blades should be autoclavable

14. Standard and security of ISO7376 standard; Manufacturer / supplier should have ISO certificate for quality standard. The lithium battery should comply to IEC 62133 or its equivalent

General Conditions

15. Each unit shall be supplied with an operational manual and a service manual in English
16. Full graphic illustrated original technical literature in English describing the equipment offered and detailing the specifications shall be supplied with the bid
17. A detailed proforma invoice of the equipment describing the parts and accessories offered as requested by the specification together with their make, model, country of origin, unit price, quantity, total price, ref. numbers of accessories and the period of warranty offered etc. shall be compulsorily provided with the bid
18. The equipment shall be covered by a comprehensive "Parts & Labour" warranty for the period of not less than 24 calendar months from the date of successful Installation & commissioning. Such a warranty shall also include servicing and at least 4 preventive maintenances per year during the period of validity.
19. Bidder shall furnish the documentary evidence to demonstrate that the good it offers meet the following usage requirement
 - i. ISO certificate for good manufacturing practice
 - ii. CE Certificate
 - iii. Shall conform to IEC / European CE Safety Standards and have current unambiguous US FDA Approval.
20. with service manual, soft and hard copy
21. with operations manual, soft and hard copy
22. with warranty:
 - i. 2 years of parts and labor
 - ii. 2 preventive maintenance visits
23. with shipping & handling costs
24. with delivery lead time of equipment
25. CIF- YAP