A .	
Announcement	No.

V6AI- AM/KUTE-FM Radio Announcement Form Department of Youth & Civics Affairs

[] Emergency – ASAP

Director, DY&CA

DHS

Starting Date: 3/2725

Ending Date: 5/15/25

[x] Regular – 3x per day

To:

From:

Frequency:

Languages (abook all that and 1)				
Languages (check all that apply):				
[x] English [x] Yapese [x] (Outer Island			
Directions:				
All announcements should be type-will XXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXX	itten and double spac XXXXXXXXXXXX	ed. If more space is ne	eded, please attach a	n extra sheet.
PLEASE	REFER TO A	ATTACHMEN	T	
XXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXX	XXXXXXXXXX	XXXXXXXXXXX	XXXXXXXXX	XXXXXXXXX
Reviewed & Approved By:	CA, Director Julia		<u>8/2</u> Dat	8/25 te
FOR	V6AI-AM/KUTE-	FM OFFICIAL USE	ONLY	
Received By:		Date:	Tir	ne:
Acting Chief of Med	ia, Jessica P. Fas			



The State of Yap

Department of Health Services
P.O. Box 148, Colonia, Yap 96943
Federated States of Micronesia
Telephone #: 011 (691) 350-2110



Request For Quotation:

The Department of Health Service is tendering its Request for Quotation (RFQ) and will be accepting sealed bids for <u>Digital Mobile X-Ray Unit</u>, which specifications are listed below, beginning on March 27, 2025 until at 3:00 pm at local standard time on May 15, 2025:

Specifications for a Digital Mobile X-Ray Unit:

- 1. POWER SUPPLY: 110 +/- 10% VOLT AC, 60 HERTZ, SINGLE PHASE
- 2. WITH BATTERY-OPERATION
- 3. WITH FLAT PANEL DETECTOR, WIRELESS TYPE WITH TWO BATTERY AND EXTERNAL BATTERY CHARGER.
- 4. MOTORIZED
- 5. X-RAY GENERATOR:
 - Microprocessor controlled high frequency inverter type and the output of 15 KW or above.
 - It should have a digital display of mAs and kV and an electronic timer.
 - KV range: 40kV to 125kV or more.
 - mAs range: 0.32 200 mAs
 - Exposure time range: 0.001 10 s
- 6. X-RAY TUBE:
 - Dual Focal spot: 0.7 or less/1.3 mm or better
 - Rotating anode. Max 125 KV
 - Heat storage capacity of the anode: 300 KHU or better
 - Tube column rotation range should be +/- 270 degrees
- 7. CONNECTIVITY:
 - Fully network-ready
 - Compatible with the Hospital's PACS System
 - Compatible with DICOM
- 8. WITH SERVICE MANUAL, soft and hard copy
- 9. WITH OPRATIONS MANUAL, soft and hard copy
- 10. WITH INSTALLATION & COMMISSIONING
- 11. WITH WARRANTY:
 - 2 years of parts and labor
 - 2 Preventive Maintenance visits
- 12. WITH A PORTABLE, SEPARATE X-RAY ANALYZER
- 13. WITH SHIPPING & HANDLING COSTS
- 14. CIF- YAP



The State of Yap

Department of Health Services
P.O. Box 148, Colonia, Yap 96943
Federated States of Micronesia
Telephone #: 011 (691) 350-2110



All sealed bids shall be addressed to the Director of Health Services, P.O. Box 148, Colonia, Yap State 96943 and shall be labeled "Bid Request For Quotation on a Digital Mobile X-Ray unit "and shall bear the name and address of the bidder on the outside of the envelope. Bid request thus received will be publicly opened and read by the Director or her designee at 3:00 pm local time on May 15, 2025.

The Department of Health Services reserves the right to reject any and all of the bids submitted. The award of contract, if awarded, will be made to the lowest responsible bidder whose proposal complies with all the requirements prescribed, in accordance with the Yap State Bidding Regulations.

Bid announcement can also be viewed at the Yap State Government Site. http://www.yapstategov.org/

Please direct any questions or inquires to Augustine Harong or Anthony Bay at 350-2110.